



Certificate of Analysis

1.04907.0100 Potassium bromide for IR spectroscopy Uvasol®
Batch B1673907

Batch Values

Application test	passes test	
Absorbance(2g pressed disk, 0.5 cm layer thickness)(Water; 3775-3050/1760-1520 cm ⁻¹)	≤ 0.15	
Absorbance(2g pressed disk, 0.5 cm layer thickness)(CH-compounds; 3050-2750 cm ⁻¹)	≤ 0.004	
Absorbance(2g pressed disk, 0.5 cm layer thickness)(other ; 4000-400 cm ⁻¹)	≤ 0.020	
Particle Size (d10)	≤ 100	µm
Particle Size (d50)	≤ 200	µm
Particle Size (d90)	≤ 500	µm

Date of release (DD.MM.YYYY) 21.02.2019
Minimum shelf life (DD.MM.YYYY) 29.02.2024

Dr. Manuel Schaffroth
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.